

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name: TONR3 L

Catalogue no.: 984372

Lot no.: H259

Package: 20 x 20 ml

Storage Conditions: 2...8 °C

Expiry date: 2013-02

**Quality Control  
Performance test:**

Control	Result (µg N/l)	Target range (µg N/l)
Ctrl High	405.48	360 – 440
Ctrl Low	98.20	90 – 100

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 18 / 4 2012



Hannele Laine  
Quality Assurance Specialist  
Quality Regulatory and Compliance  
Clinical Diagnostics Finland

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name: TONR3 L

Catalogue no.: 984372

Lot no.: G807

Package: 20 x 20 ml

Storage Conditions: 2...8 °C


Expiry date: 2012-09

**Quality Control**  
**Performance test:**

<u>Control</u>	<u>Result (mg/l)</u>	<u>Target range (mg/l)</u>
Ctrl High	0.37	0.36 – 0.44
Ctrl Low	0.10	0.09 – 0.11

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa    /    /    2011

  
\_\_\_\_\_  
Hannele Laine  
Quality Assurance Specialist  
Quality Regulatory and Compliance  
Clinical Diagnostics Finland